

Chart 7: PAA (peracetic acid) analyzers

| INTERFERING GAS | | | | | | | |
|-----------------|-----------------|-----|-----------------|------------------|------|-----------------|-------------------|
| PAA ANALYZER | NH ₃ | CO | CO ₂ | ethyl alcohol | EtO | Cl ₂ | isopropyl alcohol |
| | | 150 | >1000 | >10 ⁴ | >500 | >500 | 6 [N] |

[N] = Negative interference